

Amendments in tender number 4108 dated 21.09.2024 for the supply and installation of fully automated five-part hematology analyzer (free of cost) on Reagent rental basis for Central Lab.

Following amendments have been made in response to the PRE-BID MEETING held on 01-10-2024, 11 A.M. in the Meeting Hall, Administrative Building, RIMS, Ranchi and the representations sent on email:

Sl No.	Specifications as per tender documents	Amendments/ corrigendum
1.	Page no. 6 Type of cell counter: It should be fully automated and have five-part WBC differential along with immature granulocyte/ECG/atypical lymphocyte count.	It should be a fully automated and have five-part WBC differential along with immature granulocyte/ large immature cells count /ECG/atypical lymphocyte count.
2.	Page no. 6 Discrete analysis mode available: CBC + DIFF, CBC + NRBC	CBC + DIFF/ CBC+NRBC/ CBC+NRBC flagging
3.	Page no. 6 Platelet Analysis must include- PLC, MPV, PDW, PCT	Platelet Analysis must include- Platelet count, MPV, PDW, PCT
4.	Page 6 WBC analysis must include- TLC, PMN # & %, LYM # & %, MN # & %, EOS # & %, Baso # & %, ECG # & % / Atypical lymphocytes, two Scatter gram.	WBC analysis must include- TLC, PMN # & %, LYM # & %, MN # & %, EOS # & %, Baso # & % / ECG # & % / Atypical lymphocytes, two Scatter grams.
5.	Page no. 7 Controls: Third party control should be provided for RBC and its indices, WBC-TLC and DLC, Platelet count including PLT-F	Third party control should be provided for RBC and its indices, WBC-TLC and DLC, Platelet count. In case of unavailability of third party control, the institute will take decision on acceptance of in house control. However, this will not be a regular thing.
6.	Page no. 8 In General Terms and Conditions (Point Number -17) If number of test/parameters or patient load increases or based upon future	If number of test/parameters or patient load increases or based upon future

Anupriya Prasad
08/10/2024

High
08/10/24

2/2
08.10.24

Prasad
08/10/24

	requirement, instrument will be upgraded with advance model with advanced model with higher throughput within stipulated time as decided by the competent authority of the institute by the bidder free of cost.	requirements to maintain a TAT of two hours for CBC from the time sample received in the lab till the readiness of reports , instrument will be upgraded with advance model with higher throughput within stipulated time as decided by the competent authority of the institute by the bidder free of cost.
7.	Page 9 In General terms and Conditions (Point Number -23) Any accessories like refrigerator (to store kit) desktop computer, barcode printer, LIS interface, laser printer (Printer speed >40 pages/min), as per instrument requirement should be provided by the bidder.	Any accessories like desktop computer, barcode printer, LIS interface, laser printer (Printer speed >40 pages/min), as per instrument requirement should be provided by the bidder.

Based on the additional suggestions during pre-bid meeting, following amendments are made:

SI No	Specifications as per tender documents	Amendments/ corrigendum
1.	Page no. 12 (Point no. 56) If the Principal manufacturer (OEM) withdraws its distributorship from the bidder during the validity of tender by issuing authorization letter to another firm, the Director has full right to accept candidature of new coming distributor subject to the new firm fulfilling all other terms and conditions of the tender. No claim whatsoever will be entertained from the previous bidder (However, if any supply	If the OEM is participating in the bid and supplying the reagents through a distributor , and it withdraws its distributorship during the validity of tender by issuing authorization letter to another firm, the Director has full right to accept candidature of new coming distributor subject to the new firm fulfilling all other terms and conditions of the tender. The new distributor will have to undergo a

Anup Prasad
08/10/2024

2/10
08.10.24

Anand Kumar
08.10.24

16/10/24

	order is pending with the previous bidder, same will be executed by the new bidder).	revised rate contract. No claim whatsoever will be entertained from the previous bidder (However, if any supply order is pending with the previous bidder, same will be executed by the new bidder). Renewal of authorization from OEM for the entire period of contract for tender will be required.
2.	Page no. 6 Certification should be by any of- ISO 13485/USFDA/CE European certified for quality standards	USFDA/ European CE certified with four digit notified body for IVD
3.	Website link for the specifications of the equipment quoted should be provided.	

**Medical Superintendent cum
Chairman, Procurement Committee
Rajendra Institute of Medical Sciences, Ranchi**

Anupa Prasad
08/10/2024

rc
08.10.24

Amit Kumar
08.10.24

Raj
08/10/24