

राजेन्द्रआयुर्विज्ञानसंस्थान

(झारखण्डसरकारका एक स्वयतशासी संस्थान)

राँची-834009(झारखण्ड)

दुरभाष: 0651-2541533, फ़ैक्स: 0651-2540629,

E-mail: rimsranchi@rediffmail.com



RAJENDRA INSTITUTE OF MEDICAL SCIENCES

(An Autonomous Institute under Govt. of Jharkhand)

Ranchi-834009(Jharkhand)

Phone: 0651-2541533, Fax: 0651-2540629,

Email : rimsranchi@rediffmail.com

पत्रांक..... DR0/2024/471 दिनांक..... 09/10/2024

स्थानीय निविदा आमंत्रण सूचना

राजेन्द्रआयुर्विज्ञान संस्थान, राँची के "Department of Oral Pathology, Microbiology and Forensic Odontology, Dental Institute" विभागमें " **Generate medical imaging datasets (MIDAS) for tobacco induced oral lesions to enhance research in health in India**" नामक शोध परियोजना चल रही है। इस शोध परियोजना के लिए "Macro lens and Ring flash" की आवश्यकता है जिसके लिए स्थानीय निविदा आमंत्रित की जा रही है।

सक्षम निविदादाताओं से अनुरोध है की मुहरबंद निविदा सात (07) दिनों के अंदर डीन (रिसर्च) के कार्यालय में जमा करें। निविदा के साथ एजेन्सी का GST Registration Certificate, PAN Card एवं तीन वर्षों का ITR संलग्न करना आवश्यक होगा एवं इस आशय का स्वयंघोषणा करना बाध्यकारी होगा कि अंकित दर से कमतर पर उनके द्वारा किसी भी सरकारी संस्थान में आपूर्ति नहीं किया गया है।

I. Macro lens and Macro Ring flash

S.no.	Requirements	Specification	Quantity
1	Macro lens	SPECIFICATIONS OF MACRO LENS It should have at least 90mm focal length (180mm equivalent) with a 14° field of view • It Should have F3.5-F22 aperture range • It Should have 7 rounded aperture blades • It Should have Optical image stabilization with 6 stops of CIPA-rated stabilization • It Should have IP53 weather sealing • It Should have 2:1 magnification, up to 4:1 magnification with teleconverters • It Should have 62mm filter size	01
2	Macro ring flash	SPECIFICATIONS OF MACRO RING FLASH • Compatible with SLR Cameras • It should have Ring Flash and Control Unit • It should have 5-Step Control: 1/16, 1/8, 1/4, 1/2, 1/1 • It should Include Six Lens Adapter Rings • It should have Color Temperature: Daylight 5600K • It should have 1 to 2 sec Recycling • It should Runs on AA Batteries	01

[Signature]
Dean (Research)
Rajendra Institute of
Medical Sciences
Ranchi

डीन (रिसर्च)
राजेन्द्रआयुर्विज्ञानसंस्थान, राँची।
9/10/24